



Certificate No.: 4752-1132695  
Procedure No.: 1102

Control No.: 600851  
Page 1 of 2

**UNIT UNDER TEST**

**Manufacturer:** FLUKE  
**Model No.:** 5606  
**Serial No.:** 62740  
**Cust. Ref. No.:**  
**Description:** FULL IMMERSION PRT  
**Date Rec'd:** 12/4/2020  
**Condition Rec'd:** GOOD

**SUBMITTED BY**

**Customer:** K&S TECHNICAL SERVICES, INC  
941 NEW LONDON ROAD  
NEWARK, DE 19711  
**P.O. #:** 10355  
**Precal:** OUT OF SPEC **Final:** IN SPEC

## CALIBRATION CERTIFICATE

All calibrations are performed by qualified personnel using instrumentation, procedures and methods which guarantee specifications claimed are reliable. When specified, all calibrations are performed in accordance with current ISO/IEC 17025, ANSI/NCSL Z-540-1, MIL-STD-45662A, and ENV/Pro-Lab Quality Manual - Rev 5. Standards used are traceable to The National Institute of Standards and Technology (NIST). Expanded uncertainties are calculated using methods described in the Guide to the Expression of Uncertainty of Measurement (GUM) utilizing a coverage factor of K=2 (95% confidence) and kept on file at Pro-Lab. At a minimum, standards are selected with an uncertainty of 25% or better, where possible. This certificate and/or data shall not be reproduced except in full, without the written permission of Pro-Lab Management.

**Standards Used**

| Asset # | Description                           | Certificate Number | Date Due  |
|---------|---------------------------------------|--------------------|-----------|
| 1274    | PRECISION DIGITAL THERMOMETER READOUT | 52-VEN-2774        | 4/2/2021  |
| 01840   | Platinum Resistance Thermometer       | 52-VEN-16136       | 1/24/2021 |
| 00985   | REFERENCE THERMOMETER                 | 4753-1088851       | 9/10/2021 |

Temperature: 20.5 C  
Humidity: 32.5% RH  
Approved By: Jennifer Shelly  
Laboratory Supervisor  
E-Signed 14-Dec-2020 1:17 PM

Date Tested: 14-Dec-2020  
Date Due: 14-Dec-2021  
Calibrated By: Samuel Depue  
Calibration Technician  
E-Signed 14-Dec-2020 10:04 AM



Certificate No.: 4752-1132695  
Procedure No.: 1102

Control No.: 600851  
Page 2 of 2

**UNIT UNDER TEST**

**Manufacturer:** FLUKE  
**Model No.:** 5606  
**Serial No.:** 62740  
**Cust. Ref. No.:**  
**Description:** FULL IMMERSION PRT  
**Date Rec'd:** 12/4/2020  
**Condition Rec'd:** GOOD

**SUBMITTED BY**

**Customer:** K&S TECHNICAL SERVICES, INC  
941 NEW LONDON ROAD  
NEWARK, DE 19711  
**P.O. #:** 10355  
**Precal:** OUT OF SPEC    **Final:** IN SPEC

### CALIBRATION DATA

Note: Calibration results may drift from documented values prior to calibration due date attributable to various factors. Results obtained apply to the UUT only and are reflective of conditions at the time of this test. Measurement uncertainty is not taken into account when statements of compliance with specifications are made.

**Temperature**

| Description | Standard | As Found | Final    | Min.     | Max.       |
|-------------|----------|----------|----------|----------|------------|
| °C          | -195.000 | -194.913 | -195.014 | -195.050 | -194.950 A |
| °C          | -38.000  | -37.853  | -38.013  | -38.050  | -37.950 A  |
| °C          | 0.000    | 0.147    | -0.012   | -0.050   | 0.050 A    |
| °C          | 100.000  | 100.152  | 100.008  | 99.950   | 100.050 A  |
| °C          | 150.000  | 150.131  | 150.004  | 149.950  | 150.050 A  |